Document reference no.: 080225/1 Report no.: RN-0129



EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

MC810 MA911B/2 MA811 pH / TDS / Temperature, 230V or 115V Double Junction Ph Electrode with 2 m Cable and Gel Filled Electrolyte Solution Fixed TDS Probe

Are in compliance with:

Directive 2011/65/EU Directive 2014/30/EU Directive 2014/35/EU RoHS2 EMC and its standards Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or references to the other technical specification: EN 63000:2018 Restriction of the use of certain hazardous substances (RoHS) EC Regulation 1907/2006 Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) EN 61326-1: 2013 Electrical Equipment of Measurement, Control and Laboratory use Class A - EMC requirements Part 1: General requirements and its standards: EMC requirements Part 1: General requirements and its standards: EN 61000-4-2: 2009, EMC 4 kV / 8 kV contact / air Electrostatic Discharge Immunity Test EN 61000-4-3: 2020, EMC 6 V/m (80 Mhz to 1 GHz) Radiated RF, Electromagnetic Field Immunity Test

The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.

This declaration is the responsibility of the manufacturer:

Milwaukee Electronics Kft.	Milwaukee Instruments, Inc.
Alsó-kikötő sor 11/C	(United States of America only)
H-6726 Szeged - HUNGARY	2950 Business Park Drive – Rocky
tel: +36 62 428050	Mount NC 27804 - U.S.A
fax: +36 62 428 051	tel: +1 252 443 3630
www.milwaukeeinst.com	fax: +1 252 4431937
	www.milwaukeeinstruments.com
Test Laboratory	Milwaukee Instruments, Inc.
This is the result of test that was carried out from the submitted type-samples of a product in	Test performed by: Jozsa Ervin / ETO Eng. / [Name / Title of Signatory / Signature]
conformity with the specification of the espective standards.	Declaration authorized by: Prodan Vasile / ETO Responsible /
The declaration holder has the right to fix the CE-mark for LVD on the product complying with	Issue place / Date: Nusfalău / February 25, 2022
the inspection sample.	[Place / Date in format: Month, dd, yyyy]

FO-MWCETEST r.6 - 2022.02.24