

Document reference no.: 230921/8

Report no.: RN-0058



EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

PH58

Pocket-size pH, ORP & Temperature Meter

Are in compliance with:

Directive 2011/65/EU

RoHS2

Directive 2014/30/EU

EMC and its standards

Directive 2014/35/EU

Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or references to the other technical specification:

EN 63000:2018

Restriction of the use of certain hazardous substances (RoHS)

EC Regulation 1907/2006

Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH)

EN 61326-1: 2013

Electrical Equipment of Measurement, Control and Laboratory use

Class A

- EMC requirements Part 1: General requirements and its standards:

EMC requirements Part 1:

General requirements and its standards:

EN 61000-4-2: 2009, EMC

Electrostatic Discharge Immunity Test

4 kV / 8 kV contact / air

EN 61000-4-3: 2020, EMC

Radiated RF, Electromagnetic Field Immunity Test

10 V/m (80 Mhz to 1 GHz)

EN61010-1: 2010

Safety Requirements for Electrical Equipment for Measurement, Control and

Laboratory Use Part 1: General Requirements.

The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.

This declaration is the responsibility of the manufacturer:

Milwaukee Electronics Kft.

Alsó-kikötő sor 11/C

H-6726 Szeged - HUNGARY

tel: +36 62 428050

fax: +36 62 428 051

www.milwaukeeinst.com

Milwaukee Instruments, Inc.

(United States of America only)

2950 Business Park Drive - Rocky

Mount NC 27804 - U.S.A

tel: +1 252 443 3630

fax: +1 252 4431937

www.milwaukeeinstruments.com



Test Laboratory

This is the result of test that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.

The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample.

Milwaukee Instruments, Inc.

Test performed by: Jozsa Ervin / ETO Eng. /

[Name / Title of Signatory / Signature]

Declaration authorized by: Prodan Vasile / ETO Responsible /

[Name / Title of Signatory / Signature]

Issue place / Date: Nuşfalău / September 23, 2021

[Place / Date in format: Month, dd, yyyy]