

## EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

<b>MW805</b>	pH / EC / TDS / Temperature Portable Meter
<b>MW806</b>	pH / EC / TDS / Temperature Portable Meter
<b>MA852</b>	pH/EC/TDS and temperature Probe

Are in compliance with:	Directive 2011/65/EU	RoHS2
	Directive 2014/30/EU	EMC and its standards
	Directive 2014/35/EU	Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or references to the other technical specification:

EN 63000:2018	Restriction of the use of certain hazardous substances (RoHS)	
EC Regulation 1907/2006	Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH)	
EN 61326-1: 2013	Electrical Equipment of Measurement, Control and Laboratory use	Class A
	- EMC requirements Part 1: General requirements and its standards:	
EMC requirements Part 1:	General requirements and its standards:	
EN 61000-4-2: 2009, EMC	Electrostatic Discharge Immunity Test	4kV / 8 kV contact / air
EN 61000-4-3: 2020, EMC	Radiated RF, Electromagnetic Field Immunity Test	6V/m (80 Mhz to 1 GHz)
EN61010-1: 2010	Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use Part 1: General Requirements.	

The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.

This declaration is the responsibility of the manufacturer:

### Milwaukee Electronics Kft.

Alsó-kikötő sor 11/C  
H-6726 Szeged - HUNGARY  
tel: +36 62 428050  
fax: +36 62 428 051  
www.milwaukeeeinst.com

### Milwaukee Instruments, Inc.

(United States of America only)  
2950 Business Park Drive - Rocky  
Mount NC 27804 - U.S.A  
tel: +1 252 443 3630  
fax: +1 252 4431937  
www.milwaukeeinstruments.com



### Test Laboratory

This is the result of test that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.

The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample.

### Milwaukee Instruments, Inc.

Test performed by: Jozsa Ervin / ETO Eng. /  
[Name / Title of Signatory / Signature]

Declaration authorized by: Prodan Vasile / ETO Responsible /  
[Name / Title of Signatory / Signature]

Issue place / Date: Nuşfalău / November 25, 2021  
[Place / Date in format: Month, dd, yyyy]